## Instructions for use

# Tango XXL

**Shower-toilet chair** 





lm	_		+-	_	4.
ш	υų	"	ιd	ш	ı:

Read these instructions carefully before using the Tango XXL!

Date first used
Jser
Name
Address

#### Manufacturer

Lopital Nederland B.V. Laarakkerweg 9 5061 JR Oisterwijk The Netherlands T +31 (0)13 5239300 F +31 (0)13 5239301 www.lopital.com info@lopital.nl

Notes on symbols		
[ji]	Read the instructions for use before use.	
<u> </u>	Note.	

## **Contents**

1.	Introduction	17
1.1.	Introduction	17
1.2.	Safety	17
1.3.	Warranty	18
2.	Use	18
2.1.	First use	18
2.2.	Working with the Tango XXL	19
3.	Operation	20
3.1.	Wheel blocking	20
3.2.	Directional wheel	20
3.3.	Footrest	20
3.4.	Armrests	21
3.5.	Toilet function	21
3.5.1.	In combination with toilet	21
3.5.2.	Use of a bedpan or bucket	21
4.	Cleaning and maintenance	22
5.	Storage and transport	23
6.	Removal of parts	24
7.	Technical Specifications	24
7.1.	Structure	24
7.2.	Dimensions	24

### 1. Introduction

#### 1.1. Introduction

The shower and toilet chair Tango XXL is a medical device for caring for and washing clients in an ergonomically responsible manner. Comfort is the primary consideration for both the client and the carer.

The Tango XXL can also be used as a toilet chair, in combination with a bedpan / bucket. The Tango XXL is a shower / toilet chair with independent, folding armrests, polyurethane seat, rails for a bedpan, rotating and foldable footrests. These functions are essential performance.

The average technical life of the Tango XXL is 10 years, assuming it is cleaned daily and regularly maintained according to the maintenance schedule supplied (*See: Chapter 4 Cleaning and Maintenance*).

The Tango XXL is a class 1 product with CE-marking. This marking corresponds with directive 93/42/EC and the revised directive 2007/47/EC for medical devices of 21 March 2010.

#### 1.2. Safety



A failure to comply with the following safety points and the further definitions in these instructions for use could lead to hazardous situations. Please read carefully before use.

Use

- $\boldsymbol{\cdot}$  Make sure that the Tango XXL brake is applied before transferring clients.
- $\boldsymbol{\cdot}$  Notify the client before moving the Tango XXL.
- Make sure that the client always places his / her arms on the appropriate armrest in order to avoid anything becoming jammed.
- · Keep the armrests down / closed when moving the Tango XXL.
- Always try to move backwards over obstacles (rather than forwards).
- It is not permitted to charge the Tango XXL with more than 300 kg / 661 lbs.
- The Tango XXL may not be used to transport loads other than people.
- Only use original parts, supplied by Lopital.
- Use the Tango XXL in a clean and tidy working environment.
- · Only verifiably authorised personnel is allowed to work with the Tango XXL.
- Swimming pool locations are not considered to be 'normal conditions' and will substantially shorten the lifespan of certain components.
- $\cdot\,$  During use of the Tango XXL, the water temperature may not exceed 43°C / 109°F.

#### Maintenance

- The Tango XXL must be inspected and serviced at least once a year.
- Servicing and maintenance of the Tango XXL may only be carried out by persons authorised by Lopital.
- Constructive components of the Tango XXL may only be replaced by suitably authorised Lopital personnel.
- If erroneous use, transport, an accident or improper maintenance leads to deformation of the Tango XXL, it may no longer be used and you must contact the supplier.

In the event of breakdowns, you must contact Lopital or your local supplier. E-mail: service@lopital.nl



The Tango XXL shower-toilet chair must only be operated by persons who have been sufficiently trained and expert personnel who have experience working with the target group.



The Tango XXL must be used in a clinical setting.



Making changes to the construction could affect the safety of the Tango XXL. This will also lead to Lopital's liability and warranty conditions being invalidated. As a result, the Tango XXL will no longer comply with the guidelines for medical devices.

#### 1.3. Warranty

Lopital products are guaranteed for 2 years with respect to construction and / or material faults. In the event of varying warranty terms, the exact period will be indicated on your invoice; this also serves as the warranty itself.

Components that are subject to wear and tear will not fall under warranty unless there are clear construction and / or material faults.

#### 2. Use

#### 2.1. First use

Before you use the Tango XXL, you must check to ensure the following components are present. These components can be replaced by authorised individuals in the event of damage. These components are the applied parts of the Tango XXL. These components may come into contact with the user(s) while the Tango XXL is in use.



#### 2.2. Working with the Tango XXL

#### Step 1

If necessary, turn the armrests downwards (See paragraph 3.4. Armrests).

#### Step 2

If necessary, turn the footrests away (See paragraph 3.3 Foot supports).

#### Step 3

The Tango XXL must be put on the brakes before transferring persons (See paragraph 3.1. Wheel blocking).

#### Step 4

The client can take a seat in the chair. Support the client if necessary.

#### Step 5

Place both feet on the footrest (See paragraph 3.2 Foot supports).

#### Step 6

Close the armrests once the client has taken a seat in the chair. (See paragraph 3.4. Armrests).

#### Step 7

Notify the client about the processes you will be carrying out.

#### Step 8

Take the chair off the brake when you wish to move it (See paragraph 3.1. Wheel blocking).



Figure 1 Swivel wheel.

## 3. Operation

#### 3.1. Wheel blocking

- The brake can be applied on the Tango XXL by pushing the blue cap (1) on the wheel downwards (*See figure 1*).
- To take the Tango XXL off the brake, the lip (2) on the wheel must be pushed downwards.




Figure 2 Directional wheel.

#### 3.2. Directional wheel

- The standard Tango XXL comes with a swivel wheel that is fitted to the left front side of the device (See figure 2). The direction of the wheel can be blocked by pushing the grey cap (1) on the wheel downwards. This makes it easier to push the device over long distances as the chair is less inclined to move sideways.
- In order to remove the blocking, the lip (2) on the wheel must be pushed downwards.

#### 3.3. Footrest

- Before the client takes a seat in the Tango XXL, the footrests must be raised slightly and turned outwards. There is now sufficient room in front of the chair and the client can easily take a seat.
- Once the person has taken a seat, the carer or client raises one leg, the carer then turns the footrest forwards until it falls into the lock. The client can then rest their foot on the support.
- The same procedure is then carried out with the other leg.



Figure 3 Foot support.



Figure 4 Turning out footrest.



Always replace this component on the chair when you have finished using it. This will avoid the component being lost.

#### 3.4. Armrests

- The independent, turning armrests are easy to use.
- The safety catch is pulled outwards to unlock the armrest (1) (*See figure 5*).
- The armrest can now easily be turned to the back (2) (See figures 5 & 6).
- Always protect the client when folding the armrests up or down
- The armrests can also be used as a support when taking a seat or standing up.



Figure 5 Armrest.



Figure 6 Armrest turned out.



When folding the armrest away, avoid the hinging point so as to avoid becoming jammed.



When folding the armrests away, avoid fingers or body parts becoming jammed between the armrest and the back support.



The armrest or front closure must not be used as a lever arm (e.g. when moving the chair over an obstacle) or as a push / pull bar. You must not sit on it either. This could damage the Tango XXL and / or present a hazard.

#### 3.5. Toilet function

3.5.1. In combination with toilet

• The Tango XXL can be moved over the toilet.



#### The Tango XXL does not fit over every toilet.

3.5.2. Use of a bedpan or bucket

- If the Tango XXL is being used as a toilet chair, the special guides under the seat can be used to accommodate a stainless steel / plastic bedpan or bucket.
- The bedpan or bucket is slid in from the back of the chair (See figure 7).
- If you are using the Lopital plastic bedpan, you must slide it as far forwards as possible (See figure 8).



Figure 7 Bedpan.



Figure 8 Lopital plastic bedpan.

• If you are using a stainless steel bedpan or bucket, the bedpan can be placed in 2 positions under the seat. At the front of the seat (See figure 9) or at the back of the seat (See figure 10).



In order to avoid injuries, we advise the carer to remove the lid from the seat and insert the bucket or bedpan before the client takes a seat on the chair.



Figure 9 Stainless steel bedpan, front.



The Lopital plastic bedpan, stainless steel bedpan and bucket must be cleaned and disinfected after use, for example in a bedpan washing unit.



If the Lopital plastic bedpan, stainless steel bedpan or bucket has been used, it must be removed from the chair before the chair is moved or adjusted in height.



Figure 10 Stainless steel bedpan, back.

## 4. Cleaning and maintenance

The Tango XXL must be cleaned after every use, with water and regular household cleaning agents. Remove excess soap residues. PUR integrated foam components with IMC coating should be rubbed clean and disinfected. Do not use any abrasive or other aggressive cleaning products (e.g. chlorine bleach). Practical experience has shown that regular household cleaning agents with limited alcohol content (max. 5%) are most suitable for this purpose. The product Neoquat combi® is recommended for disinfection purposes (on the basis of long-term testing). If necessary, the PUR components can be cleaned with a solution of 70% ethanol or propanol.

You must ensure that both the disinfecting and cleaning processes involve the components being rubbed and not just submerged. The time the products must be left on the components has a significant impact upon the lifespan of the PUR components. As a result of the material properties of PUR, contact with sharp objects must be avoided. This could lead to damage or tears in the material. Finally, we would like to point out that non-critical parts of the PUR components must also be cleaned and decontaminated.

Check the wheels regularly for hairs, dust and soap residues and remove when necessary. Maintenance and repairs to the Tango XXL may only be carried out by authorised personnel.

Maintenance schedule			
After every use	Clean chair with water and regular household cleaning products.		
Daily	<ul> <li>PUR foam components must be cleaned and disinfected.</li> <li>Check to ensure all parts of the Tango XXL are present, including the footrests and head support.</li> </ul>		
Weekly	<ul> <li>Clean frame.</li> <li>Check wheels for hairs, dust and soap residues and remove if necessary.</li> <li>Check that the chair is still working correctly.</li> </ul>		
Annually  lopital  2017  2017  Service & Onderhoud Tel: +31 (0)13 523 9300  www.lopital.nl	According to the guidelines for medical devices, a chair must be maintained in line with the manufacturer's instructions. This process will preferably be carried out by Lopital or another authorised body. Lopital offers the option of having the chair thoroughly inspected and serviced once a year via Periodic Preventative Maintenance.  For more information about the maintenance contract, please contact Lopital (service@lopital.nl) or speak to your supplier.		

## 5. Storage and transport

During transport and storage, the ambient conditions must remain moderate. In other words, the temperature should remain between -40 and +55 degrees Celsius (-40 to 130 degrees Fahrenheit) and humidity levels should be kept as low as possible. Vibration during transport must be kept to a minimum in order to avoid damaging the hinged mechanisms on the chair.

## 6. Removal of parts

All parts that are replaced or removed can be sent back to Lopital BV. You may also offer the parts to the engineer for destruction. We will ensure that all parts are processed in an environmentally-friendly manner.

## 7. Technical Specifications

#### 7.1. Structure

Frame. stainless steel and steel tubes and plates that are

equipped with a 2-layer powder coating

stainless steel, cover caps made of plastic V0 Hinge points:

plastic, Ø 12.5 cm / 4.7 inch, double blocking facility Wheels:

Body support: PUR (polyurethane). 30.5 kg / 67 lbs Weight: 300 kg / 661 lbs Maximum permitted load:

#### 7.2. Dimensions

Surface area: 83 cm wide / 32.7 inch, 104 cm / 40.9 inch deep

Seat height: 51 cm / 20.1 inch

